

Continuous Radon Monitor

Model Number: 1028
Calibration Date: 03/21/2014
Monitor Time: 11/10/2014 16:42

Serial Number: 55540044
CF: 2.79

Inspection Company

TEK Inspections
P.O. Box 33368
Reno, NV-89533
Phone Number: 775-746-4990
License Number:

Billing Information

123 Your Street
Anywhere, NV-12345

Site Information

123 Your Street
Anywhere, NV-12345

Site & Condition


Wind: NA
Year Built: NA
Mitigation System: Not Installed
SqFt: NA

Atmospheric Condition: NA
Structure Type: NA
Monitor Location: NA

Test Summary

Start Time: 11/07/2014 11:01
End Time: 11/10/2014 16:01
Measurement Interval(hr): 1.0
Exposure Time: 3 Days 5 hrs

Overall Avg: 2.2 pCi/l
EPA Avg: 2.2 pCi/l



***** 11/07/2014 *****

Time	Counts pCi/l	Flags
12:01	2.9	
13:01	1.1	
14:01	1.4	
15:01	1.1	
16:01	1.4	
17:01	2.1	
18:01	2.9	
19:01	0.7	
20:01	0.7	
21:01	2.1	
22:01	1.1	
23:01	0.7	

***** 11/08/2014 *****

Time	Counts pCi/l	Flags
00:01	1.1	
01:01	1.1	
02:01	1.8	
03:01	2.5	
04:01	1.8	
05:01	0.4	
06:01	2.1	
07:01	1.1	
08:01	0.4	
09:01	1.8	
10:01	1.4	
11:01	1.8	
12:01	0.4	
13:01	0.7	
14:01	0.4	
15:01	4.3	
16:01	2.9	
17:01	1.4	
18:01	0.7	
19:01	1.8	
20:01	2.5	
21:01	3.6	
22:01	2.5	
23:01	1.4	

***** 11/09/2014 *****

Time	Counts pCi/l	Flags
00:01	3.6	
01:01	3.9	

***** 11/09/2014 *****

Time	Counts pCi/l	Flags
02:01	2.1	
03:01	3.6	
04:01	2.5	
05:01	1.4	
06:01	2.1	
07:01	1.4	
08:01	3.2	
09:01	2.9	
10:01	1.8	
11:01	2.1	
12:01	2.9	
13:01	3.2	
14:01	4.7	
15:01	5.0	
16:01	1.8	
17:01	3.2	
18:01	4.7	
19:01	4.7	
20:01	3.9	
21:01	3.6	
22:01	3.6	
23:01	1.8	

***** 11/10/2014 *****

Time	Counts pCi/l	Flags
00:01	1.8	
01:01	1.1	
02:01	3.6	
03:01	1.4	
04:01	3.6	
05:01	1.8	
06:01	1.8	
07:01	4.3	
08:01	1.4	
09:01	2.9	
10:01	2.9	
11:01	2.5	
12:01	1.4	
13:01	0.4	
14:01	1.1	
15:01	1.4	
16:01	2.5	

Error Flags:
M Motion:

Inspector Signature _____

PC Software Version: 2.2.0
Embedded Software Version: 109